

Medical Teaching Institution (MTI)

Khyber Teaching Hospital (KTH)

Peshawar

Contact No: 091-9224400

BID SOLICITATION DOCUMENTS

For

CSSD Consumables 2025-2026

Note: The prospective bidder is expected to examine the Bidding Documents carefully, including all Instructions, Terms & Conditions, and Specifications, etc. Failure to furnish all information required by the Bidding documents or submission of a Bid not substantially responsive to the Bidding Documents in every respect would result in the rejection of the Bid.

INTRODUCTION:

Medical Teaching Institution (MTI)/ Khyber Teaching Hospital Peshawar invites. Sealed bids From the eligible bidders for the procurement of CSSD Consumables for Hospital under Open Competitive Bidding under rule 6(2) (a) "**Single Stage Single Envelope**" bidding procedures of Khyber Pakhtunkhwa Public Procurement Regulatory Authority (KPPRA) Rules 2014.

1) INSTRUCTIONS TO BIDDERS:

1. This Bidding procedure will be conducted in light of Khyber Pakhtunkhwa Public Procurement Regulatory Authority (KPPRA) Laws and rules made there under, along with Standard Bidding Documents.
2. The bids will be opened on 05-3-25 in the presence of the bidders/representatives who choose to attend
3. Any bid received after the deadline for submission of bids shall not be entertained and shall be returned unopened to the Bidder.
4. The bid should be complete in all respects and must be signed by the bidder.
5. All prices quoted must be in Pak Rupees (PKR) and should include all applicable taxes. If not specifically mentioned in the Quotation, it will be presumed that the prices include all the taxes. GST Shall be included where applicable
6. Bidders are essentially required to provide correct and latest postal/email/web addresses, and phone/mobile/fax numbers for active and timely communication.
7. For any query, or clarification regarding Services / Bid Solicitation Documents, the applicants may send a written request at least one day prior to the opening date.
8. The Bidder may after its submission withdraw its bid prior to the expiry of the deadline prescribed for submission of bids. Withdrawn bids will be returned unopened to the Bidders.
9. Any bid not received as per the terms and conditions laid down in this document are liable to be ignored. No offer shall be considered if:
 - a. Received without earnest money;
 - b. It is received after the date and time fixed for its receipt;
 - c. The offer is ambiguous;
 - d. The offer is conditional i.e. advance payment, or currency fluctuations etc.;
 - e. The offer is from blacklisted firm in any Federal / Provincial Govt. Deptt:
 - f. Only typed tender on original prescribing letter pad, sealed & signed should be submitted, the quoted Price must be pre printed and hand written quoted price will not be acceptable. The tenders must be according to hospital specification; alternate rates (**Double rates for single Items**) will not be accepted.
10. Bids will be rejected if the Bid is in some way connected with bids submitted under names different from his own.

11. In case of Bid Tie, the decision will be taken by making toss/draw/Recall sealed quotation in front of the bidders.
12. Any direct or indirect effort by a bidding firm to influence this institution during the process of selection of a bidder or award of contract may besides rejection of its bid result into its disqualification from participation in the KTH Peshawar's future bids.
13. The brand name must be mentioned on each quoted item.

2) ELIGIBILITY CRITERIA:

- Bidders must give compliance to the below mentioned clauses as these are mandatory to being eligible for the bidding process. Relevant certificates must be attached.
- The bidder must be registered with Income / Sales Tax Department, reflected as Active Tax Payer on the list of FBR.NTN and for services KNTN.
- The bidder shall provide an undertaking that the bidder has not been declared black listed by any Governmental/ Semi-Governmental institutions.
- The bidder shall provide an undertaking that supplies will be according to approved sample throughout the year, if in any case low standard or quality compromised found in supplies, that supply will be confiscated along with legal action against the bidder.
- Sample submission within 1 week after sample submission letter issued to each bidder, if samples not provided on specific time period the bid will be considered as non-responsive
- Bidders shall not be eligible to bid if they are under a declaration of Ineligibility for corrupt and fraudulent practices issued by any government organization in accordance with the Section 44(1) KPP Rules 2014.
- The bidder must be manufacturer/ Importer / authorization certificate.
- **Local office (Peshawar)**

3. GENERAL CONDITIONS: -

1. KTH Peshawar shall evaluate the bid in a manner prescribed in advance, without reference to the price and reject any proposal which does not conform to the specified requirements.
2. Alternative bid shall not be considered and shall be rejected by the Competent Authority.
3. At any time prior to the deadline for submission of bids, KTH Peshawar may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective Bidder, modify the bidding documents by amendment.
4. If a bid is not substantially responsive, it will be rejected by the Procuring Entity and may not subsequently be made responsive by the Bidder by correction of the nonconformity.
5. KTH MTI Peshawar may accept or reject any or all of the bids under Rule 47 of KPPRA Rules, 2014.

**Khyber Teaching Hospital
Medical Teaching institute
Peshawar**

Phone: 091-9224400

INVITATION FOR BIDS

Hospital Director, Medical Teaching Institute, **Khyber Teaching Hospital Peshawar**

Invites sealed tenders under National Competitive Bidding for the procurement of CSSD Consumables for Hospital, under rule 6(2)(a) ***“single stage Single envelope procedure”*** of KPPRA Rules 2014, from Income / Sales tax, reflected on Active Taxpayer List of FBR.

The bidders are required to submit bid security Rs-250,000/- in shape of Call Deposit Receipt (CDR) in the name of Hospital Director KTH Peshawar.

The tenders complete in all respects must reach the undersigned by 10:00 hrs. on 05-3-25, which will be opened at 10:30 hrs on the same day in the office of the Manager Supply Chain in the presence of the tender opening committee and the bidders/representatives who may choose to attend.

Competent Authority reserves the right to reject any or all the bids as per provisions contained in Rule 47 of KPPRA Rules 2014.

4. BID Security

Bid security Rs-**250,000/-** in shape of Call Deposit refundable drawn in favor of "Hospital Director KTH Hospital"

The bid security may be forfeited:

- i) If a Bidder withdraws its bid during the period of bid validity specified by the Bidder on the Bid Form; or
- ii) In the case of a successful Bidder, if the Bidder fails to sign the contract
- iii) If a successful Bidder unable to supply the items after issuance of 3 reminders, order will be placed to the next lowest bidder.

5. BID VALIDITY:

- i) The bids should be valid for a period of 90 Days.
- ii) In exceptional circumstances, KTH Hospital may solicit the Bidder's consent to an extension of the period of validity reasons shall be recorded in writing. The request and the responses there to shall be made in writing. The bid security provided shall also be suitably extended. A Bidder may refuse the request without forfeiting its bid security. A Bidder granting the request will not be required nor permitted to modify its bid, except as provided in the bidding document.

6. BID VALIDITY:

Rates approved will be valid till 30-06-2026.

PROPOSED CSSD CONSUMABLES LIST

Item #	Item Name	U/M	Qty
PACKING PAPER (CREPE PAPER)			
1	Crepe Paper Sheet Green (50x50) cm x 500 sheets CE/FDA Approved	Pack/Case	
2	Crepe Paper Sheet Green (60x60) cm x 500 sheets CE/FDA Approved	Pack/Case	
3	Crepe Paper Green (75x75) cm x 250 sheets CE/FDA Approved	Pack/Case	
4	Crepe Paper Green (90x90) cm x 250 sheets CE/FDA Approved	Pack/Case	
5	Crepe Paper Green (100x100) cm x 250 sheets CE/FDA Approved	Pack/Case	
6	Crepe Paper Green (120x120) cm x 100 sheets CE/FDA Approved	Pack/Case	
PACKING REELS (FLAT & GUSSETED)			
7	STERILE PACKAGING REEL, FLAT WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, 3" (7,5 X 200 MTR), medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
8	STERILE PACKAGING REEL, FLAT WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, 4" (10 X 200 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
9	STERILE PACKAGING REEL, FLAT WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, 6" (15 X 200 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
10	STERILE PACKAGING REEL WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, FLAT, 8" (20 X 200 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
11	STERILE PACKAGING REEL WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, FLAT, 10" (25 X 200 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
12	STERILE PACKAGING REEL WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, FLAT, 12" (30 X 200 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	

13	STERILE PACKAGING REEL WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, FLAT, 4" (10 X 200 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
14	STERILE PACKAGING REEL, GUSSE WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, 6" (15 X 100 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
15	STERILE PACKAGING REEL, GUSSE WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, 8" (20 X 100 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS). Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
16	STERILE PACKAGING REEL, GUSSE WITH BUILT IN PROCESS INDICATORS FOR STEAM, EO,FO STERILIZATION, 10" (25 X 100 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS).Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
17	STERILE PACKAGING REEL, GUSSE WITH BUILT IN PROCESS INDICATORS FOR STEAM,EO,FO STERILIZATION, 12" (30 X 100 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS).Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
18	STERILE PACKAGING REEL, GUSSE WITH BUILT IN PROCESS INDICATORS FOR STEAM,EO,FO STERILIZATION, 14" (35 X 100 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS).Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
19	STERILE PACKAGING REEL, GUSSE WITH BUILT IN PROCESS INDICATORS FOR STEAM,EO,FO STERILIZATION, 16" (40 X 100 MTR) medical grade paper 60gms.: Film thickness 52 micron (FILM PET 12 / PP 40 SHATTERLESS).Reel have a shelf-life of 5 years with test reports. CE/FDA Approved	Roll	
20	STERILE PACKAGING REEL Tyvek, 4" x 70 Mtr. CE/FDA Approved	Roll	
21	STERILE PACKAGING REEL Tyvek, 6" x 70 Mtr. CE/FDA Approved	Roll	
22	STERILE PACKAGING REEL Tyvek, 8" x 70 Mtr. CE/FDA Approved	Roll	
23	STERILE PACKAGING REEL Tyvek, 10" x 70 Mtr. CE/FDA Approved	Roll	
24	STERILE PACKAGING REEL Tyvek, 12" x 70 Mtr. CE/FDA Approved	Roll	
Heat Sealing Pouches			
	Flat pouch size 150*300mm (CE/FDA Approved) 1000/Pack		

	Flat pouch size 250*350mm (CE/FDA Approved) 1000/Pack		
	Flat pouch size 250*400mm (CE/FDA Approved) 1000/Pack		
	Flat pouch size 300*400mm (CE/FDA Approved) 1000/Pack		
	Flat pouch size 300*450mm (CE/FDA Approved) 1000/Pack		

Sterilization Monitoring, Validation & Documentation Products

25	Lead Free & latex Free Autoclave tape for Steam. 18MM x 50 Mtr.	Roll	
26	Bowie Dick test Start-up kit, consisting of a Compact-PCD® and 100 BDS-adhesive indicator strips. Simulating 7 kg cotton pack according to EN ISO 11140-4.	Set	
27	Batch Monitoring Start-up kit, consisting of 1 Compact-PCD for solid, porous & Hollow loads testing according to EN 867-5 and 100 adhesive indicator strips.	Set	
28	Refill Pack with adhesive indicator strips for BDS (Box of 100 strips)	Box	
29	Refill Pack with adhesive indicators strips for BMS PCD (Box of 500 Strips)	Box	
30	Manual Printing 3 Line Labelling Gun	Nos.	
31	Double adhesive 3 Line labels with process indicator for Steam according to EN ISO 11140-1.	Roll of 750	
32	Double adhesive 3 Line labels with process indicator for H2O2 according to EN ISO 11140-1.	Roll of 750	
33	Self-contained Biological Indicators for steam	Box of 100.	
34	Self-Contained Biological indicator for H2O2	Box of 50.	
35	Bio Compact PCD for Biological Indicators according to EN 867-5	Nos.	
36	Incubator with aluminum Block for variable temperature selection and programming of the incubation time	Nos.	
37	Self-Adhesive Package Monitoring Indicators for steam Type 4 according to EN ISO 11140-1.	Box of 3200 Strips	
38	Package Monitoring Indicators for H2O2 Box of 3200	Box	
39	Refill packs with Level 2 self-adhesive cleaning indicators for validation and monitoring of cleaning processes. (single indicators)	Box of 320	
40	Cleaning indicator holder stainless steel to be used with Cleaning Indicators for WDs.	Nos.	

MANUAL/AUTOMATIC CLEANING /WASHING & DISINFECTION SOLUTIONS

41	Mildly alkaline/enzymatic instrument Cleaner i.e. combination of surfactants & sophisticated enzymes, biodegradable & doesn't require neutralization. Cleaning Temperatures should be 35-45 C° .	Can of 5 Ltr.	
42	Neutral Rinsing and drying aid with slight Lubricant effect. approx. PH 7.9	Can of 5 Ltr.	
43	Low / intermediate level Cleaner & Disinfectant for Manual/Ultrasonic cleaning & Disinfection of instruments (1 component product). Should be VAH (German Association for Applied Hygiene). Aldehyde free. Service life should be 14 Days.	Btl. of 1 Ltr.	

44	Descaler/troubleshooter for Stainless steel instruments with Passivation characteristics. approx. pH 2.2-1.1	Can of 5 Ltr.	
45	Manual Lubricant for treatment of surgical instruments, CFC & silicon Free. According to European Pharmacopia Recommendations.	Bottle of 400 ml	
46	High Level disinfectant Solution, ready to use with prepared solution service life of 14 days with validation strips (paracetic acid solution)	Can of 5 Ltr.	
47	Surface Disinfection Solution, aldehyde free		
48	Hydrogen peroxide in aqueous solution compatible to PLASMA Sterilizer Machine.	Nos.	
49	Reusable Jar / Bottle Brush	NOS.	
50	Autoclave Regular Brush Cleaning Kit	NOS	
51	Asymmetric Double wire Brush	NOS	
52	Broncho brushes	NOS	
53	Nylon twisted wire brushes	NOS	
54	Frazier/Baron suction tubes	NOS	
55	Frazier/Baron suction tubes	NOS	
56	Nylon twisted wire brushes	NOS	
57	Reusable External Cleaning Brush	NOS	
58	20" / 50 cm Asymmetric Double-End Wire Brush	NOS	
59	26" / 65 cm Asymmetric Double-End Wire Brush	NOS	
60	16" / 406 mm Nylon Twisted Wire Brushes	NOS	
61	Instruments Cleaning Brushes Kit 1-Rigid reusable flat brush 3- Flexible reusable flat brush Different sizes 4- Extra Flexible reusable flat brush	NOS	
62	Instrument Identification Tape	NOS	
63	Water Filters	NOS	
64	RO Membranes	NOS	
65	Mix Bed Resin	NOS	
66	Pure white Crystal Salt	NOS	
67	Heat Protective Gloves for Unloading of sterile Goods	Pairs	
68	Protection Gloves for Decontamination area	Pairs	
69	Protection Shoe Covers for decontamination Area	Nos.	
70	Cartridges for shoe cover Machine	Nos.	
71	Instruments foam gel spray	Nos.	
72	SS cleaner	Nos.	
73	Silicon Spray for autoclave gasket	Nos.	
74	Thermal Paper for Autoclave for CISA -Matachana	Nos	
75	Cassette for plasma Sterilizer	Nos	

Note: Samples may be called for all the above items if required.

7. Award of Contract:

Contracts shall be confirmed through a written agreement signed by the successful bidder and the MTI /KTH Peshawar.

8. Payment:

- a. No advance payment will be permissible.
- b. The payment will be made after a successful supply and inspection