







KHYBER TEACHING HOSPITAL MTI PESHAWAR

Technical Evaluation Pathology Lab Equipments on Regent Rental Basis FY-2023-26

Firm and Product Evaluation Total Marks =70		Hematology Analyzer 5 Part					Coagulation	
Sr.	Parametre	Marks	HOORA	Abbott	NexTelc	5.Ejazuddin	Barhani	
1	Compliance to Purchaser's Specifications	Fully compliance to the required specifications as per statement of Requirement (Up to a maximum of four Minor deviations may be accommodated subject to the condition that main function and performance of the equipment in any aspect would not be adversely affected, however, one mark will be deducted for each minor deviation, (Up to 4 marks will be deducted)	30	30	30	30	30	
2	Vis-à-vis Format	Vis-à-vis/Compliance sheet will not be considered as a technical Quotation	Mandatory	yes	yes	Yes	yes	
3	Product Certification	FDA	4	4	4	4	0	
		CE	2	2	2	0	2	
		MHLW	2	0	0	0	2	
		Out of the above three certificates, ONE certificate is mandatory						
4	Quality Certificates	Valid ISO 13485 Quality management certificate	2	2	0	2	2	
5	Product's Performance	Satisfactory Performance Certificates for the quoted model from the medical institutions within Pakistan (Supply order /Purchase order will not be considered as a per performance certificate) Each certificate carries two mark. Must be on the Institution letter head duly signed and stamped by HOD/Directors.	10	2	6	4	2	
6	Firm Evaluation & Legal Requirements	Manufacturer Authorization Certificate, or Partnership Deed with manufacturer (original certificate to be produced during technical evaluation)	Mandatory	yes	yes	Yes	yes	
		Satisfactory performance of the firm from medical institution in KPK, (Supply order/Purchase order will not be considered as a performance certificate) Each certificate carries two marks. Must be on the Institution letter head duly signed and stamped by HOD/Directors.	8	6	6	0	0	6
		Satisfactory performance certificate(s) issued by well reputable organizations/KTH each certificate carries one mark. (National wide) Must be on the Institution letter head duly signed and stamped by HOD/Directors	6	6	2	2	0	0
			5	5	5	5	5	5
7	Technical Staff	Manufacturer trained engineer of the quoted equipment based in Peshawar. (To be verified with relevant documents/certificates)			5			
		Manufacturer trained service engineer of the quoted equipment at national level (To be verified with relevant documents/Certificates)	Mandatory	yes	yes	Yes	yes	yes
8	Supplier's office for msintenance and 24/7 support	Availability of workshop at National level	Mandatory	yes	yes	Yes	yes	yes
		Certificate to the affect that the firm will provide 95% uptime of quoted equipment	Mandatory	yes	yes	Yes	yes	yes
		Certificate for provision of response time of 2 Hours or less to a service call (during working hours 08am to 05pm)	Mandatory	yes	yes	Yes	yes	yes
		Certificate for provision of onsite training in the use of equipment to the relevant staff.	Mandatory	yes	yes	Yes	yes	yes
		List of tools, testing equipment and calibration equipment relevant to the product	1	1	1	0	1	1
Total Marks			70	58	56	51	50	50


Chairman Pathology Deptt-KTH

Rep. of Finance-KTH


Incharge Medical ICU-KTH

Manager Supply Chain-KTH


Manager Biomedical Engineering-KTH

Chairman Technical Evaluation Committee KTH

KHYBER TEACHING HOSPITAL MTI PESHAWAR

Technical Evaluation Pathology Lab Equipments on Regent Rental Basis FY-2023-26

Sr.	Parametre	Sub parameter	Marks	ESR		Plasmapheresis
				Vantage	Sind Medical Store	Parcon Distribution
1	Compliance to Purchaser's Specifications	Fully compliance to the required specifications as per statement of Requirement (Up to a maximum of four Minor deviations may be accommodated subject to the condition that main function and performance of the equipment in any aspect would not be adversely affected, however, one mark will be deducted for each minor deviation, (Up to 4 marks will be deducted).	30	30	30	26
2	Vis-à-vis Format	Vis-à-vis/Compliance sheet will not be considered as a technical Quotation	Mandatory	yes	yes	yes
3	Product Certification	FDA	4	0	0	
		CE	2	2	2	2
		MHLW	2			
4	Quality Certificates	Out of the above three certificates, ONE certificate is mandatory Valid ISO 13485 Quality management certificate	2	2	2	0
5	Product's Performance	Satisfactory Performance Certificates for the quoted model from the medical institutions within Pakistan.(Supply order /Purchase order will not be considered as a per performance certificate) Each certificate carries two mark. Must be on the Institution letter head duly signed and stamped by HOD/Directors.	10	8	0	0
6	Firm Evaluation & Legal Requirements	Manufacturer Authorization Certificate, or Partnership Deed with manufacturer (original certificate to be produced during technical evaluation)	Mandatory	yes	yes	yes
		Satisfactory performance of the firm from medical institution in kpk, (Supply order/Purchase order will not be considered as a performance certificate) Each certificate carries two marks. Must be on the Institution letter head duly signed and stamped by HOD/Directors	8	2	2	0
		Satisfactory performance certificate(s) issued by well reputable organizations/KTH each certificate carries one mark. (National wide) Must be on the Institution letter head duly signed and stamped by HOD/Directors	6	6	6	0
7	Technical Staff	Manufacturer trained engineer of the quoted equipment based in Peshawar. (To be verified with relevant documents/certificates)	5	5	0	0
		Manufacturer trained service engineer of the quoted equipment at national level. (To be verified with relevant documents/Certificates)	Mandatory	yes	yes	not provided
8	Supplier's office for maintenance and 24/7 support	Availability of workshop at National level	Mandatory	yes	yes	yes
		Certificate to the affect that the firm will provide 95% uptime of quoted equipment	Mandatory	yes	yes	yes
		Certificate for provision of response time of 2 Hours or less to a service call (during working hours 08am to 05pm)	Mandatory	yes	yes	yes
		Certificate for provision of onsite training in the use of equipment to the relevant staff.	Mandatory	yes	yes	yes
		List of tools, testing equipment and calibration equipment relevant to the product	1	1	0	1
Total Marks			70	56	42	29

Chairman Pathology Deptt-KTH

Rep. of Finance-KTH

Incharge Medical ICU-KTH

Manager Supply Chain-KTH

Manager Biomedical Engineering-KTH

Chairman Technical Evaluation Committee KTH